1/2/2018

SOL # 252

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Serial No.: ARWF-0394

Calibration File No.:	01631	Calib. Date:	09/15/2017	Calib. No.:	00038
Certification File No.:	01632	Cert. Date:	09/15/2017	Cert. No.:	00028
Linearity File No.:	01633	Lin. Date:	09/15/2017	Lin. No.:	00028
Solution File No.:	01655	Soln. Date:	01/02/2018	Soln. No.:	00252
Sequential File No.:	01655	File Date:	01/02/2018		

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110
Model No.: CU-34
Serial No.: DDAN-0040
Expires: 03/20/2019
Bottle No.: 0212

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:09S		
Control 1 EC	0.102%	12:09S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:10S		
Control 2 EC	0.099%	12:11S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12S		
Control 3 EC	0.100%	12:12S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: SCHWARTZ First Name: STEVE MI: C
Signature: _____ Badge No.: 1293
Date: 01/02/2018