

Soln # 262
1/14/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02147
Certification File No.: 02148
Linearity File No.: 02149
Solution File No.: 02244
Sequential File No.: 02244

Serial No.: ARAJ-0075

Calib. Date: 09/15/2017
Cert. Date: 09/15/2017
Lin. Date: 09/15/2017
Soln. Date: 01/14/2018
File Date: 01/14/2018

Calib. No.: 00029
Cert. No.: 00020
Lin. No.: 00020
Soln. No.: 00262

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 09/14/2019
Bottle No.: 1372

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:55S		
Control 1 EC	0.098%	01:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:56S		
Control 2 EC	0.098%	01:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:57S		
Control 3 EC	0.098%	01:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: *P. Gambardella #365*

Badge No.: 365

Date: 01/14/2018