

SOLN 264
1/30/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02255
Certification File No.: 02256
Linearity File No.: 02257
Solution File No.: 02258
Sequential File No.: 02258

Serial No.: ARAJ-0075
Calib. No.: 00030
Cert. No.: 00021
Lin. No.: 00021
Soln. No.: 00264
File Date: 01/30/2018

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34
Serial No.: DDAN-0041
Expires: 03/20/2019
Bottle No.: 0504

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:44S		
Control 1 EC	0.102%	09:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		
Control 2 EC	0.100%	09:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47S		
Control 3 EC	0.100%	09:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: JXKPZ-436 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TPT. I R W. #6704

Badge No.: 6704

Date: 01/30/2018