

Calibrating Unit New Standard Solution Report



ORIGINAL

SOLN # 267

2/25/18

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 02255
 Certification File No.: 02256
 Linearity File No.: 02257
 Solution File No.: 02284
 Sequential File No.: 02284

Calib. Date: 01/30/2018
 Cert. Date: 01/30/2018
 Lin. Date: 01/30/2018
 Soln. Date: 02/25/2018
 File Date: 02/25/2018

Serial No.: ARAJ-0075
 Calib. No.: 00030
 Cert. No.: 00021
 Lin. No.: 00021
 Soln. No.: 00267

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0041
 Expires: 09/14/2019
 Bottle No.: 1493

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:31S		
Control 1 EC	0.099%	08:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:33S		
Control 2 EC	0.099%	08:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		
Control 3 EC	0.099%	08:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

W Job

Badge No.: 306

Date: 02/25/2018