

SOLN# 268

3/8/18

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02255  
Certification File No.: 02256  
Linearity File No.: 02257  
Solution File No.: 02296  
Sequential File No.: 02296

Calib. Date: 01/30/2018  
Cert. Date: 01/30/2018  
Lin. Date: 01/30/2018  
Soln. Date: 03/08/2018  
File Date: 03/08/2018

Serial No.: ARAJ-0075  
Calib. No.: 00030  
Cert. No.: 00021  
Lin. No.: 00021  
Soln. No.: 00268

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 09/14/2019  
Bottle No.: 1359

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:15S		
Control 1 EC	0.099%	14:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:17S		
Control 2 EC	0.099%	14:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18S		
Control 3 EC	0.100%	14:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:20S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/08/2018