

SOLN: 270

3/23/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02255
Certification File No.: 02256
Linearity File No.: 02257
Solution File No.: 02310
Sequential File No.: 02310

Serial No.: ARAJ-0075

Calib. Date: 01/30/2018
Cert. Date: 01/30/2018
Lin. Date: 01/30/2018
Soln. Date: 03/23/2018
File Date: 03/23/2018

Calib. No.: 00030
Cert. No.: 00021
Lin. No.: 00021
Soln. No.: 00270

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 09/14/2019
Bottle No.: 1492

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:25S		
Control 1 EC	0.100%	14:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27S		
Control 2 EC	0.098%	14:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28S		
Control 3 EC	0.098%	14:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: *Pascal Gambardella* #365

Badge No.: 365

Date: 03/23/2018