

SOLN # 271
4-6-2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	01/30/2018
Calibration File No.:	02255	Calib. No.:	00030
Certification File No.:	02256	Cert. Date:	01/30/2018
Linearity File No.:	02257	Lin. Date:	01/30/2018
Solution File No.:	02318	Soln. Date:	04/06/2018
Sequential File No.:	02318	File Date:	04/06/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1300

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:57D		
Control 1 EC	0.100%	11:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 2 EC	0.099%	11:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		
Control 3 EC	0.099%	12:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:

 #365

Badge No.: 365

Date: 04/06/2018