

SOLN\* 272

4/14/18

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02255	Calib. Date:	01/30/2018
Certification File No.:	02256	Cert. No.:	00030
Linearity File No.:	02257	Cert. Date:	01/30/2018
Solution File No.:	02323	Lin. No.:	00021
Sequential File No.:	02323	Lin. Date:	01/30/2018
		Soln. Date:	04/14/2018
		Soln. No.:	00272
		File Date:	04/14/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	0605


Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:59D		
Control 1 EC	0.100%	14:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		
Control 2 EC	0.098%	14:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 3 EC	0.099%	14:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KOEPPEN	First Name: ADAM	MI: M
--------------------	------------------	-------

Signature:  #350	Badge No.: 350
	Date: 04/14/2018