

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01659 Calib. Date: 01/30/2018 Calib. No.: 00039  
Certification File No.: 01660 Cert. Date: 01/30/2018 Cert. No.: 00029  
Linearity File No.: 01661 Lin. Date: 01/30/2018 Lin. No.: 00029  
Solution File No.: 01671 Soln. Date: 04/20/2018 Soln. No.: 00258  
Sequential File No.: 01671 File Date: 04/20/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
Control Solution %: 0.100% Expires: 09/14/2019  
Solution Control Lot: 17290 Bottle No.: 1369

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:56D		
Control 1 EC	0.101%	12:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:57D		
Control 2 EC	0.100%	12:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		
Control 3 EC	0.100%	13:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 04/20/2018