

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 02255  
 Certification File No.: 02256  
 Linearity File No.: 02257  
 Solution File No.: 02335  
 Sequential File No.: 02335

Serial No.: ARAJ-0075

Calib. Date: 01/30/2018      Calib. No.: 00030  
 Cert. Date: 01/30/2018      Cert. No.: 00021  
 Lin. Date: 01/30/2018      Lin. No.: 00021  
 Soln. Date: 04/27/2018      Soln. No.: 00273  
 File Date: 04/27/2018

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 17290

Model No.: CU-34      Serial No.: DDAN-0041  
 Expires: 09/14/2019  
 Bottle No.: 1482

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:45D		
Control 1 EC	0.098%	13:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47D		
Control 2 EC	0.097%	13:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		
Control 3 EC	0.097%	13:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_

*Pascal J. [Signature]* #365

Badge No.: 365

Date: 04/27/2018