

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02255
Certification File No.: 02256
Linearity File No.: 02257
Solution File No.: 02340
Sequential File No.: 02340

Serial No.: ARAJ-0075

Calib. Date: 01/30/2018
Cert. Date: 01/30/2018
Lin. Date: 01/30/2018
Soln. Date: 05/01/2018
File Date: 05/01/2018

Calib. No.: 00030
Cert. No.: 00021
Lin. No.: 00021
Soln. No.: 00274

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 09/14/2019
Bottle No.: 1286

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:44D		
Control 1 EC	0.097%	11:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 2 EC	0.097%	11:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:47D		
Control 3 EC	0.097%	11:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: _____



Badge No.: 336

Date: 05/01/2018