

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02255	Calib. Date: 01/30/2018
Certification File No.:	02256	Calib. No.: 00030
Linearity File No.:	02257	Cert. Date: 01/30/2018
Solution File No.:	02345	Cert. No.: 00021
Sequential File No.:	02345	Lin. Date: 01/30/2018
		Lin. No.: 00021
		Soln. Date: 05/06/2018
		Soln. No.: 00275
		File Date: 05/06/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 1186

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:28D		
Control 1 EC	0.099%	03:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:29D		
Control 2 EC	0.098%	03:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:31D		
Control 3 EC	0.098%	03:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____



Badge No.: 365

Date: 05/06/2018