

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01659 Calib. Date: 01/30/2018 Calib. No.: 00039
Certification File No.: 01660 Cert. Date: 01/30/2018 Cert. No.: 00029
Linearity File No.: 01661 Lin. Date: 01/30/2018 Lin. No.: 00029
Solution File No.: 01677 Soln. Date: 05/11/2018 Soln. No.: 00259
Sequential File No.: 01677 File Date: 05/11/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 1242

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:47D		
Control 1 EC	0.100%	13:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		
Control 2 EC	0.099%	13:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		
Control 3 EC	0.100%	13:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 05/11/2018