

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01659	Calib. Date: 01/30/2018
Certification File No.:	01660	Calib. No.: 00039
Linearity File No.:	01661	Cert. Date: 01/30/2018
Solution File No.:	01678	Cert. No.: 00029
Sequential File No.:	01678	Lin. Date: 01/30/2018
		Lin. No.: 00029
		Soln. Date: 05/23/2018
		Soln. No.: 00260
		File Date: 05/23/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 1340

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:12D		
Control 1 EC	0.100%	13:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		
Control 2 EC	0.100%	13:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		
Control 3 EC	0.100%	13:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:16D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 05/23/2018