

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02255
Certification File No.: 02256
Linearity File No.: 02257
Solution File No.: 02358
Sequential File No.: 02358

Serial No.: ARAJ-0075

Calib. Date: 01/30/2018 Calib. No.: 00030
Cert. Date: 01/30/2018 Cert. No.: 00021
Lin. Date: 01/30/2018 Lin. No.: 00021
Soln. Date: 05/23/2018 Soln. No.: 00276
File Date: 05/23/2018

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17420

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 12/11/2019
Bottle No.: 1163

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:10D		
Control 1 EC	0.101%	15:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:11D		
Control 2 EC	0.100%	15:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12D		
Control 3 EC	0.100%	15:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  # 350

Badge No.: 350

Date: 05/23/2018