

SOLN 279
6/28/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02255
Certification File No.: 02256
Linearity File No.: 02257
Solution File No.: 02386
Sequential File No.: 02386

Serial No.: ARAJ-0075
Calib. Date: 01/30/2018 Calib. No.: 00030
Cert. Date: 01/30/2018 Cert. No.: 00021
Lin. Date: 01/30/2018 Lin. No.: 00021
Soln. Date: 06/28/2018 Soln. No.: 00279
File Date: 06/28/2018

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17420

Model No.: CU-34
Serial No.: DDAN-0041
Expires: 12/11/2019
Bottle No.: 1137

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:38D		
Control 1 EC	0.101%	13:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		
Control 2 EC	0.101%	13:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40D		
Control 3 EC	0.100%	13:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  # 350

Badge No.: 350

Date: 06/28/2018