

SOLN 281
7/18/18

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARAJ-0075 |
| Location: | TOMS RIVER P.D. | | |
| Calibration File No.: | 02397 | Calib. Date: | 07/18/2018 |
| Certification File No.: | 02398 | Cert. Date: | 07/18/2018 |
| Linearity File No.: | 02399 | Lin. Date: | 07/18/2018 |
| Solution File No.: | 02400 | Soln. Date: | 07/18/2018 |
| Sequential File No.: | 02400 | File Date: | 07/18/2018 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWK S3-0386 |
| Solution Control Lot: | 17420 | Expires: | 12/11/2019 |
| | | Bottle No.: | 1157 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 10:03D | | |
| Control 1 EC | 0.102% | 10:03D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 10:03D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:04D | | |
| Control 2 EC | 0.101% | 10:05D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 10:05D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:05D | | |
| Control 3 EC | 0.100% | 10:06D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 10:06D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:07D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2 - 025 (KA)

Changed By:
Last Name: ALCOTT First Name: KEVIN MI: W.

Signature: Sgt. K W. Q #6704 Badge No.: 6704 Date: 07/18/2018