

SOLW 282

7/30/18

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. No.:	00031
Calibration File No.:	02397	Cert. Date:	07/18/2018
Certification File No.:	02398	Lin. Date:	07/18/2018
Linearity File No.:	02399	Soln. Date:	07/30/2018
Solution File No.:	02411	File Date:	07/30/2018
Sequential File No.:	02411		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1141

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:27D		
Control 1 EC	0.099%	08:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:28D		
Control 2 EC	0.098%	08:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:29D		
Control 3 EC	0.097%	08:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 07/30/2018