

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.	Calib. Date:	07/18/2018
Calibration File No.:	01682	Calib. No.:	00040
Certification File No.:	01683	Cert. Date:	07/18/2018
Linearity File No.:	01684	Lin. Date:	07/18/2018
Solution File No.:	01689	Soln. Date:	08/10/2018
Sequential File No.:	01689	File Date:	08/10/2018

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0387
Control Solution %:	0.100%			Expires:	12/11/2019
Solution Control Lot:	17420			Bottle No.:	1166

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:09D		
Control 1 EC	0.100%	19:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:10D		
Control 2 EC	0.099%	19:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:12D		
Control 3 EC	0.099%	19:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:13D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 08/10/2018