

SOLN 284

8/19/18



ORIGINAL New Standard Solution Report

Calibrating Unit

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
 Location: TOMS RIVER P.D.
 Calibration File No.: 02397 Calib. Date: 07/18/2018 Calib. No.: 00031
 Certification File No.: 02398 Cert. Date: 07/18/2018 Cert. No.: 00022
 Linearity File No.: 02399 Lin. Date: 07/18/2018 Lin. No.: 00022
 Solution File No.: 02433 Soln. Date: 08/19/2018 Soln. No.: 00284
 Sequential File No.: 02433 File Date: 08/19/2018


Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
 Control Solution %: 0.100% Expires: 12/11/2019
 Solution Control Lot: 17420 Bottle No.: 1139

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:23D		
Control 1 EC	0.098%	12:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		
Control 2 EC	0.097%	12:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		
Control 3 EC	0.098%	12:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTHALL First Name: ROBERT MI: S
 Signature:  Badge No.: 306
 Date: 08/19/2018