

SOLN 286
9-14-18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02397
Certification File No.: 02398
Linearity File No.: 02399
Solution File No.: 02457
Sequential File No.: 02457

Serial No.: ARAJ-0075
Calib. Date: 07/18/2018
Cert. Date: 07/18/2018
Lin. Date: 07/18/2018
Soln. Date: 09/14/2018
File Date: 09/14/2018

Calib. No.: 00031
Cert. No.: 00022
Lin. No.: 00022
Soln. No.: 00286

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17420

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 12/11/2019
Bottle No.: 1146

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:10D		
Control 1 EC	0.102%	12:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.101%	12:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		
Control 3 EC	0.101%	12:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 09/14/2018