

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01682	Calib. Date: 07/18/2018
Certification File No.:	01683	Calib. No.: 00040
Linearity File No.:	01684	Cert. Date: 07/18/2018
Solution File No.:	01699	Cert. No.: 00030
Sequential File No.:	01699	Lin. Date: 07/18/2018
		Lin. No.: 00030
		Soln. Date: 09/21/2018
		Soln. No.: 00267
		File Date: 09/21/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 1147

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:22D		
Control 1 EC	0.101%	12:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23D		
Control 2 EC	0.101%	12:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24D		
Control 3 EC	0.101%	12:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Badge No.: 365

Date: 09/21/2018