

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01682	Calib. Date: 07/18/2018
Certification File No.:	01683	Calib. No.: 00040
Linearity File No.:	01684	Cert. Date: 07/18/2018
Solution File No.:	01701	Cert. No.: 00030
Sequential File No.:	01701	Lin. Date: 07/18/2018
		Lin. No.: 00030
		Soln. Date: 11/02/2018
		Soln. No.: 00269
		File Date: 11/02/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 1149

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:29S		
Control 1 EC	0.102%	13:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:30S		
Control 2 EC	0.101%	13:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31S		
Control 3 EC	0.101%	13:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_



Badge No.: 365

Date: 11/02/2018