

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02397	Calib. Date: 07/18/2018
Certification File No.:	02398	Calib. No.: 00031
Linearity File No.:	02399	Cert. Date: 07/18/2018
Solution File No.:	02522	Cert. No.: 00022
Sequential File No.:	02522	Lin. Date: 07/18/2018
		Lin. No.: 00022
		Soln. Date: 12/14/2018
		Soln. No.: 00294
		File Date: 12/14/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1364

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:51S		
Control 1 EC	0.102%	08:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52S		
Control 2 EC	0.100%	08:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		
Control 3 EC	0.099%	08:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA	First Name: PASCAL	MI: M
Signature: <u> #365</u>	Badge No.: 365	Date: 12/14/2018