ITEM 14 SOLN 272
12/27/18

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.				
Calibration File No.:	01714	Calib. Date:	12/27/2018	Calib. No.:	00041
Certification File No.:	01715	Cert. Date:	12/27/2018	Cert. No.:	00031
Linearity File No.:	01716	Lin. Date:	12/27/2018	Lin. No.:	00031
Solution File No.:	01717	Soln. Date:	12/27/2018	Soln. No.:	00272
Sequential File No.:	01717	File Date:	12/27/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0387
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	1422

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:21S		
Control 1 EC	0.101%	15:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22S		
Control 2 EC	0.100%	15:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24S		
Control 3 EC	0.100%	15:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-026 (DN)

Changed By:

Last Name: NAPOLITANO First Name: DAVID MI: M.

Signature: 7237 Badge No.: 7237

Date: 12/27/2018