

ITEM 14 SOLN 295

12/27/18

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02535	Calib. Date:	12/27/2018
Certification File No.:	02536	Cert. Date:	12/27/2018
Linearity File No.:	02537	Lin. Date:	12/27/2018
Solution File No.:	02538	Soln. Date:	12/27/2018
Sequential File No.:	02538	File Date:	12/27/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	1429

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:37S		
Control 1 EC	0.100%	10:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39S		
Control 2 EC	0.097%	10:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40S		
Control 3 EC	0.097%	10:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-025 (DN)

Changed By:
Last Name: NAPOLITANO First Name: DAVID MI: M.

Signature: *David Napolitano* 7237 Badge No.: 7237 Date: 12/27/2018