

SOLN 297
12/29/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02535
Certification File No.: 02536
Linearity File No.: 02537
Solution File No.: 02542
Sequential File No.: 02542

Serial No.: ARAJ-0075

Calib. Date: 12/27/2018
Cert. Date: 12/27/2018
Lin. Date: 12/27/2018
Soln. Date: 12/29/2018
File Date: 12/29/2018

Calib. No.: 00032
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00297

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 18150

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 05/10/2020
Bottle No.: 1185

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:00S		
Control 1 EC	0.096%	21:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:01S		
Control 2 EC	0.096%	21:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03S		
Control 3 EC	0.096%	21:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 12/29/2018