

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02535
Certification File No.: 02536
Linearity File No.: 02537
Solution File No.: 02560
Sequential File No.: 02560

Serial No.: ARAJ-0075

Calib. Date: 12/27/2018 **Calib. No.:** 00032
Cert. Date: 12/27/2018 **Cert. No.:** 00023
Lin. Date: 12/27/2018 **Lin. No.:** 00023
Soln. Date: 01/20/2019 **Soln. No.:** 00302
File Date: 01/20/2019

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 18150

Model No.: CU-34 **Serial No.:** DDWK S3-0386
Expires: 05/10/2020
Bottle No.: 1308

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:24S		
Control 1 EC	0.098%	13:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:26S		
Control 2 EC	0.097%	13:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27S		
Control 3 EC	0.095%	13:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: R. S. WMP # J⁶

Badge No.: 306

Date: 01/20/2019