

SOLN 303

1/25/19


Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02535	Calib. Date:	12/27/2018
Certification File No.:	02536	Cert. No.:	00032
Linearity File No.:	02537	Cert. Date:	12/27/2018
Solution File No.:	02562	Lin. Date:	12/27/2018
Sequential File No.:	02562	Lin. No.:	00023
		Soln. Date:	01/25/2019
		Soln. No.:	00303
		File Date:	01/25/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	1310

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:33S		
Control 1 EC	0.099%	11:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35S		
Control 2 EC	0.097%	11:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:36S		
Control 3 EC	0.096%	11:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: GAMBARDELLA First Name: PASCAL MI: M
 Signature:  #365
 Badge No.: 365
 Date: 01/25/2019