

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02535	Calib. Date: 12/27/2018
Certification File No.:	02536	Calib. No.: 00032
Linearity File No.:	02537	Cert. Date: 12/27/2018
Solution File No.:	02567	Cert. No.: 00023
Sequential File No.:	02567	Lin. Date: 12/27/2018
		Lin. No.: 00023
		Soln. Date: 02/01/2019
		Soln. No.: 00304
		File Date: 02/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1254

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:04S		
Control 1 EC	0.097%	12:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		
Control 2 EC	0.096%	12:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07S		
Control 3 EC	0.095%	12:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Badge No.: 365

Date: 02/01/2019