

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01714	Calib. Date: 12/27/2018
Certification File No.:	01715	Calib. No.: 00041
Linearity File No.:	01716	Cert. Date: 12/27/2018
Solution File No.:	01729	Cert. No.: 00031
Sequential File No.:	01729	Lin. Date: 12/27/2018
		Lin. No.: 00031
		Soln. Date: 02/02/2019
		Soln. No.: 00274
		File Date: 02/02/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1309

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:05S		
Control 1 EC	0.100%	09:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		
Control 2 EC	0.097%	09:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09S		
Control 3 EC	0.097%	09:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WORTH

First Name: ADAM

MI: C

Signature: _____



Badge No.: 378

Date: 02/02/2019