

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02535  
Certification File No.: 02536  
Linearity File No.: 02537  
Solution File No.: 02573  
Sequential File No.: 02573

Calib. Date: 12/27/2018  
Cert. Date: 12/27/2018  
Lin. Date: 12/27/2018  
Soln. Date: 02/07/2019  
File Date: 02/07/2019

Serial No.: ARAJ-0075

Calib. No.: 00032  
Cert. No.: 00023  
Lin. No.: 00023  
Soln. No.: 00306

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 18320

Model No.: CU-34

Serial No.: DDWK S3-0386  
Expires: 09/26/2020  
Bottle No.: 0923

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:25S		
Control 1 EC	0.101%	12:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27S		
Control 2 EC	0.099%	12:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28S		
Control 3 EC	0.097%	12:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 02/07/2019