

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02535	Calib. Date: 12/27/2018
Certification File No.:	02536	Calib. No.: 00032
Linearity File No.:	02537	Cert. Date: 12/27/2018
Solution File No.:	02577	Cert. No.: 00023
Sequential File No.:	02577	Lin. Date: 12/27/2018
		Lin. No.: 00023
		Soln. Date: 02/16/2019
		Soln. No.: 00307
		File Date: 02/16/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1304

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:55S		
Control 1 EC	0.100%	09:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57S		
Control 2 EC	0.099%	09:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 3 EC	0.097%	09:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BAYER

First Name: RONALD

MI: C

Signature: _____



Badge No.: 1355

Date: 02/16/2019