

Calibrating Unit New Standard Solution Report



ORIGINAL

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02535 Calib. Date: 12/27/2018 Calib. No.: 00032
Certification File No.: 02536 Cert. Date: 12/27/2018 Cert. No.: 00023
Linearity File No.: 02537 Lin. Date: 12/27/2018 Lin. No.: 00023
Solution File No.: 02594 Soln. Date: 03/09/2019 Soln. No.: 00311
Sequential File No.: 02594 File Date: 03/09/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 1187

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:39S		
Control 1 EC	0.099%	21:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40S		
Control 2 EC	0.097%	21:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		
Control 3 EC	0.096%	21:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 03/09/2019