

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 02535 Calib. Date: 12/27/2018 Calib. No.: 00032  
Certification File No.: 02536 Cert. Date: 12/27/2018 Cert. No.: 00023  
Linearity File No.: 02537 Lin. Date: 12/27/2018 Lin. No.: 00023  
Solution File No.: 02598 Soln. Date: 03/18/2019 Soln. No.: 00313  
Sequential File No.: 02598 File Date: 03/18/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
Control Solution %: 0.100% Expires: 09/26/2020  
Solution Control Lot: 18320 Bottle No.: 1147

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:51S		
Control 1 EC	0.098%	19:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:52S		
Control 2 EC	0.097%	19:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54S		
Control 3 EC	0.096%	19:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: PH. J. [Signature] #336

Badge No.: 336

Date: 03/18/2019