

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02535	Calib. Date: 12/27/2018
Certification File No.:	02536	Calib. No.: 00032
Linearity File No.:	02537	Cert. Date: 12/27/2018
Solution File No.:	02602	Cert. No.: 00023
Sequential File No.:	02602	Lin. Date: 12/27/2018
		Lin. No.: 00023
		Soln. Date: 03/23/2019
		Soln. No.: 00314
		File Date: 03/23/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0895

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:14S		
Control 1 EC	0.098%	21:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15S		
Control 2 EC	0.097%	21:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:17S		
Control 3 EC	0.096%	21:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:18S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/23/2019