

Calibrating Unit New Standard Solution Report



ORIGINAL

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02535
Certification File No.: 02536
Linearity File No.: 02537
Solution File No.: 02606
Sequential File No.: 02606

Calib. Date: 12/27/2018
Cert. Date: 12/27/2018
Lin. Date: 12/27/2018
Soln. Date: 03/26/2019
File Date: 03/26/2019

Serial No.: ARAJ-0075
Calib. No.: 00032
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00315

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 18320

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 09/26/2020
Bottle No.: 0949

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:40S		
Control 1 EC	0.100%	07:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41S		
Control 2 EC	0.098%	07:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		
Control 3 EC	0.096%	07:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

J06

Badge No.: 306

Date: 03/26/2019