

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-----------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWF-0394 |
| Location: | TOMS RIVER P.D. | |
| Calibration File No.: | 01714 | Calib. No.: 00041 |
| Certification File No.: | 01715 | Cert. No.: 00031 |
| Linearity File No.: | 01716 | Lin. No.: 00031 |
| Solution File No.: | 01743 | Soln. No.: 00279 |
| Sequential File No.: | 01743 | File Date: 03/28/2019 |

| | | | |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 | Serial No.: DDWK S3-0387 |
| Control Solution %: | 0.100% | | Expires: 09/26/2020 |
| Solution Control Lot: | 18320 | | Bottle No.: 1149 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 13:08S | | |
| Control 1 EC | -.---% | 13:08S | 24.4°C | SIMULATOR TEMP. ERROR |
| Control 1 IR | -.---% | 13:08S | 24.4°C | SIMULATOR TEMP. ERROR |
| Ambient Air Blank | -.---% | --:-- | | |
| Control 2 EC | -.---% | --:-- | N/A | |
| Control 2 IR | -.---% | --:-- | N/A | |
| Ambient Air Blank | -.---% | --:-- | | |
| Control 3 EC | -.---% | --:-- | N/A | |
| Control 3 IR | -.---% | --:-- | N/A | |
| Ambient Air Blank | -.---% | --:-- | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/28/2019