

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS-RIVER P.D.	
Calibration File No.:	01714	Calib. Date: 12/27/2018
Certification File No.:	01715	Calib. No.: 00041
Linearity File No.:	01716	Cert. Date: 12/27/2018
Solution File No.:	01744	Cert. No.: 00031
Sequential File No.:	01744	Lin. Date: 12/27/2018
		Lin. No.: 00031
		Soln. Date: 03/28/2019
		Soln. No.: 00280
		File Date: 03/28/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 1149

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:00S		
Control 1 EC	0.100%	17:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:02S		
Control 2 EC	0.099%	17:02S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03S		
Control 3 EC	0.099%	17:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Pascal Gambardella #365

Badge No.: 365

Date: 03/28/2019