

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02535	Calib. Date:	12/27/2018
Certification File No.:	02536	Cert. No.:	00032
Linearity File No.:	02537	Cert. Date:	12/27/2018
Solution File No.:	02608	Lin. Date:	12/27/2018
Sequential File No.:	02608	Lin. No.:	00023
		Soln. Date:	03/30/2019
		Soln. No.:	00316
		File Date:	03/30/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	18320	Expires:	09/26/2020
		Bottle No.:	0926

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:31S		
Control 1 EC	0.099%	10:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32S		
Control 2 EC	0.097%	10:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33S		
Control 3 EC	0.095%	10:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/30/2019