

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01714 Calib. Date: 12/27/2018 Calib. No.: 00041
Certification File No.: 01715 Cert. Date: 12/27/2018 Cert. No.: 00031
Linearity File No.: 01716 Lin. Date: 12/27/2018 Lin. No.: 00031
Solution File No.: 01750 Soln. Date: 04/06/2019 Soln. No.: 00281
Sequential File No.: 01750 File Date: 04/06/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 09/26/2020
Solution Control Lot: 18320 Bottle No.: 0896

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:41S		
Control 1 EC	0.100%	10:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42S		
Control 2 EC	0.097%	10:43S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44S		
Control 3 EC	0.098%	10:45S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 04/06/2019