

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01714	Calib. Date:	12/27/2018
Certification File No.:	01715	Cert. No.:	00041
Linearity File No.:	01716	Cert. Date:	12/27/2018
Solution File No.:	01766	Lin. No.:	00031
Sequential File No.:	01766	Lin. Date:	12/27/2018
		Soln. Date:	04/17/2019
		Soln. No.:	00283
		File Date:	04/17/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	18320	Expires:	09/26/2020
		Bottle No.:	1086

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:12D		
Control 1 EC	0.000%	17:13D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.000%	17:13D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:13D		
Control 2 EC	----%	17:14D	33.9°C	CTRL GAS SUPPLY
Control 2 IR	----%	17:14D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: 

Badge No.: 336

Date: 04/17/2019

