

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02535	Calib. Date: 12/27/2018
Certification File No.:	02536	Calib. No.: 00032
Linearity File No.:	02537	Cert. Date: 12/27/2018
Solution File No.:	02614	Cert. No.: 00023
Sequential File No.:	02614	Lin. Date: 12/27/2018
		Lin. No.: 00023
		Soln. Date: 04/21/2019
		Soln. No.: 00318
		File Date: 04/21/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 1045

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:23D		
Control 1 EC	0.099%	00:23D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24D		
Control 2 EC	0.096%	00:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:25D		
Control 3 EC	0.096%	00:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: _____

Adam Koepen #350

Badge No.: 350

Date: 04/21/2019