

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02535	Calib. Date: 12/27/2018
Certification File No.:	02536	Calib. No.: 00032
Linearity File No.:	02537	Cert. Date: 12/27/2018
Solution File No.:	02615	Cert. No.: 00023
Sequential File No.:	02615	Lin. Date: 12/27/2018
		Lin. No.: 00023
		Soln. Date: 04/26/2019
		Soln. No.: 00319
		File Date: 04/26/2019

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWK S3-0386
Control Solution %: 0.100%		Expires: 09/26/2020
Solution Control Lot: 18320		Bottle No.: 0893

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:17D		
Control 1 EC	0.097%	12:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	12:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19D		
Control 2 EC	0.096%	12:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	12:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		
Control 3 EC	0.096%	12:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	12:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 04/26/2019