

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01767 Calib. Date: 05/01/2019 Calib. No.: 00042  
 Certification File No.: 01768 Cert. Date: 05/01/2019 Cert. No.: 00032  
 Linearity File No.: 01769 Lin. Date: 05/01/2019 Lin. No.: 00032  
 Solution File No.: 01770 Soln. Date: 05/01/2019 Soln. No.: 00284  
 Sequential File No.: 01770 File Date: 05/01/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
 Control Solution %: 0.100% Expires: 09/26/2020  
 Solution Control Lot: 18320 Bottle No.: 0910

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:14D		
Control 1 EC	0.100%	10:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		
Control 2 EC	0.099%	10:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 3 EC	0.099%	10:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKPL-444 (DL)

**Changed By:**  
 Last Name: LUTZ First Name: DENNIS MI: J

Signature: Tpr I Lutz 7045 Badge No.: 7045  
 Date: 05/01/2019