

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02617	Calib. Date: 05/01/2019
Certification File No.:	02618	Calib. No.: 00033
Linearity File No.:	02619	Cert. Date: 05/01/2019
Solution File No.:	02620	Cert. No.: 00024
Sequential File No.:	02620	Lin. Date: 05/01/2019
		Lin. No.: 00024
		Soln. Date: 05/01/2019
		Soln. No.: 00320
		File Date: 05/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 1050

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:17D		
Control 1 EC	0.099%	15:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 2 EC	0.099%	15:19D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20D		
Control 3 EC	0.099%	15:20D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-436 (Di)

**Changed By:**

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: \_\_\_\_\_

*Tpr I [Signature] 7045*

Badge No.: 7045

Date: 05/01/2019