

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01767	Calib. Date: 05/01/2019
Certification File No.:	01768	Calib. No.: 00042
Linearity File No.:	01769	Cert. Date: 05/01/2019
Solution File No.:	01771	Cert. No.: 00032
Sequential File No.:	01771	Lin. Date: 05/01/2019
		Lin. No.: 00032
		Soln. Date: 05/22/2019
		Soln. No.: 00285
		File Date: 05/22/2019

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDAN-0040
Control Solution %: 0.100%		Expires: 09/26/2020
Solution Control Lot: 18320		Bottle No.: 1148

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:10D		
Control 1 EC	0.099%	17:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:11D		
Control 2 EC	0.098%	17:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:13D		
Control 3 EC	0.098%	17:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 05/22/2019