

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARAJ-0075 |
| Location: | TOMS RIVER P.D. | | |
| Calibration File No.: | 02617 | Calib. Date: | 05/01/2019 |
| Certification File No.: | 02618 | Cert. No.: | 00033 |
| Linearity File No.: | 02619 | Cert. Date: | 05/01/2019 |
| Solution File No.: | 02628 | Lin. Date: | 05/01/2019 |
| Sequential File No.: | 02628 | Lin. No.: | 00024 |
| | | Soln. Date: | 05/22/2019 |
| | | Soln. No.: | 00321 |
| | | File Date: | 05/22/2019 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDAN-0041 |
| Solution Control Lot: | 19060 | Expires: | 02/11/2021 |
| | | Bottle No.: | 0073 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 14:43D | | |
| Control 1 EC | 0.100% | 14:43D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 14:43D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:44D | | |
| Control 2 EC | 0.099% | 14:45D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 14:45D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:45D | | |
| Control 3 EC | 0.099% | 14:46D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 14:46D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:47D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA First Name: PASCAL MI: M

Signature:  #365 Badge No.: 365
 Date: 05/22/2019