

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02617 Calib. Date: 05/01/2019 Calib. No.: 00033
Certification File No.: 02618 Cert. Date: 05/01/2019 Cert. No.: 00024
Linearity File No.: 02619 Lin. Date: 05/01/2019 Lin. No.: 00024
Solution File No.: 02645 Soln. Date: 06/15/2019 Soln. No.: 00323
Sequential File No.: 02645 File Date: 06/15/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 09/26/2020
Solution Control Lot: 18320 Bottle No.: 1119

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:14D		
Control 1 EC	0.100%	04:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16D		
Control 2 EC	0.099%	04:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:17D		
Control 3 EC	0.098%	04:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 06/15/2019