

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01767 Calib. Date: 05/01/2019 Calib. No.: 00042  
Certification File No.: 01768 Cert. Date: 05/01/2019 Cert. No.: 00032  
Linearity File No.: 01769 Lin. Date: 05/01/2019 Lin. No.: 00032  
Solution File No.: 01773 Soln. Date: 06/24/2019 Soln. No.: 00286  
Sequential File No.: 01773 File Date: 06/24/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
Control Solution %: 0.100% Expires: 02/11/2021  
Solution Control Lot: 19060 Bottle No.: 0074

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:47D		
Control 1 EC	0.100%	13:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		
Control 2 EC	0.100%	13:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		
Control 3 EC	0.099%	13:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BARTOSHEK

First Name: DAVID

MI: G

Signature: 

Badge No.: 316

Date: 06/24/2019